


# Technical data sheet

OTICON | Open

CIC 75



	Oticon Open 1	Oticon Open 2	Oticon Open 3
<b>Speech Understanding</b>			
OpenSound Navigator™	Level 1	Level 2	Level 3
Max. noise removal	9 dB	5 dB	3 dB
Speech Guard™ LX	Level 1	Level 2	Level 3
Spatial Sound™ LX***	4 estimators (o)	2 estimators (o)	2 estimators (o)
Soft Speech Booster LX	•	•	•
Speech Rescue™ LX	•	•	•
<b>Sound Quality</b>			
Clear Dynamics	•	•	-
Spatial Noise Management***	o	o	-
Fitting Bandwidth*	10 KHz	8 KHz	8 KHz
Processing Channels	64	48	48
<b>Listening Comfort</b>			
Transient Noise Management	4 configurations	On/Off	On/Off
Feedback shield LX	•	•	•
Binaural Coordination****	o	o	o
YouMatic™ LX	3 configurations	2 configurations	1 configuration
<b>Personalization &amp; Optimizing Fitting</b>			
Fitting Bands	16	14	12
Adaptation Management	•	•	•
Oticon Firmware Updater	•	•	•
Fitting Formulas	VAC+, NAL-NL1+2, DSL v5.0	VAC+, NAL-NL1+2, DSL v5.0	VAC+, NAL-NL1+2, DSL v5.0
Acoustic Notifications	•	•	•
Tinnitus SoundSupport™****	o	o	o
<b>Battery life, hours**</b>	70-80	70-80	70-80

\* Bandwidth accessible for gain adjustments during fitting

\*\* Battery size 10 - IEC PR70.

Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels.

\*\*\* If NFMI is chosen

\*\*\*\* If NFMI and push button is chosen

- Default
- o Optional
- Not included

OpenSound Navigator™ continuously analyzes the environment and attenuates the disturbing noise.

NFMI wireless technology is optional and provides binaural communication

Oticon Open is built on the Velox™ platform, providing frequency resolution in 64 channels (Open 1).

Fully programmable with updatable firmware, the Velox platform is ready for the future.



IP68



Absolute Hearing Solutions LLC  
absolutehearing@att.net

750 Cross Pointe Road Suite F

Gahanna, Ohio 43230

Phone: 614-452-4280  
Toll Free: 888-803-2159  
Fax: 614-577-0461  
[www.absolutedearingsolutions.com](http://www.absolutedearingsolutions.com)

Call  
Absolute Hearing Solutions  
For Hearing Aid Details!  
614-452-4280

[www.absolutedearingsolutions.com](http://www.absolutedearingsolutions.com)


oticon  
PEOPLE FIRST

Technical data		Ear Simulator			2CC Coupler		
Measured according to		IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010			ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006		
Oticon Opn CIC 75		Opn 1	Opn 2	Opn 3	Opn 1	Opn 2	Opn 3
Frequency range Hz		100-9500	100-7500	100-7500	100-7000	100-7000	100-7000
OSPL90	Peak		119 dB SPL			109 dB SPL	
	1600 Hz		110 dB SPL			103 dB SPL	
	HFA-OSPL90		110 dB SPL			104 dB SPL	
Full-on gain*	Peak		57 dB			47 dB	
	1600 Hz		49 dB			42 dB	
	HFA-FOG		49 dB			42 dB	
Reference test gain			36 dB			27 dB	
Telecoil output (1600 Hz)	1 mA/m field		-			-	
	10 mA/m field		-			-	
	SPLITS L/R		-			-	
Total harmonic distortion (Input 70 dB SPL)	500 Hz		2 %			< 2 %	
	800 Hz		2 %			< 2 %	
	1600 Hz		3 %			2 %	
Equivalent input noise level	Omni		20 dB SPL			19 dB SPL	
Battery consumption**	Typical		1.0 mA			1.0 mA	
	Quiescent		1.0 mA			1.0 mA	
Battery life, calculated, hours***			100			100	
IRIL (IEC 60118-13:2016)		700/1400/2000 MHz: 18/12/11 dB SPL					

\* Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.

\*\* Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

\*\*\* Based on the standardised battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.



Technical information: Omnidirectional mode is used unless otherwise stated.

Operating conditions	Storage and transportation conditions
Temperature: +1°C to +40°C	Temperature and humidity should not exceed the following limits for extended periods during transportation and storage.
Relative humidity: 5% to 93%, non-condensing	Temperature: -25°C to +60°C Relative humidity: 5% to 93%, non-condensing