

Instructions for use

Call Absolute Hearing Solutions
614-452-4280



BTE SP
BTE UP

Oticon Xceed
Oticon Xceed Play

Made for
 iPhone | iPad | iPod

oticon
PEOPLE FIRST

Call Absolute Hearing Solutions
614-452-4280

Model overview

This booklet is valid for Oticon Xceed and Oticon Xceed Play in the following hearing aid models:

- ☐ BTE SP (Battery size 13)
- ☐ BTE UP (Battery size 675)

FW 8

- | | |
|---|---------------------------|
| <input type="checkbox"/> Oticon Xceed 1 | GTIN: (01) 05707131357571 |
| <input type="checkbox"/> Oticon Xceed 2 | GTIN: (01) 05707131357564 |
| <input type="checkbox"/> Oticon Xceed 3 | GTIN: (01) 05707131357557 |

FW 8

- | | |
|--|---------------------------|
| <input type="checkbox"/> Oticon Xceed Play 1 | GTIN: (01) 05707131357540 |
| <input type="checkbox"/> Oticon Xceed Play 2 | GTIN: (01) 05707131357533 |

- ☐ LED light for visual indication. Helps caregivers, parents, and teachers to operate the hearing aid and to give instructions on relevant functions and modes (see relevant information throughout the booklet).

Introduction to this booklet

This booklet gives you guidance on how to use and maintain your new hearing aid. Please read the booklet carefully, including the **Warnings** section. This will help you to get the most out of your new hearing aid.

Your hearing care professional has adjusted the hearing aid to meet your needs. If you have additional questions, please contact your hearing care professional.

| **About** | [Start-up](#) | [Handling](#) | [Options](#) | [Tinnitus](#) | [Warnings](#) | [More info](#) |

For your convenience, this booklet contains a navigation bar to help you easily navigate through the different sections.

Call Absolute Hearing Solutions
614-452-4280

Intended use

The hearing aid is intended to amplify and transmit sound to the ear and thereby compensate for impaired hearing within moderate-to-severe to profound hearing loss. The hearing aid is intended to be used by adults, children and infants.

IMPORTANT NOTICE

The hearing aid amplification is uniquely adjusted and optimized to your personal hearing capabilities during the hearing aid fitting performed by your hearing care professional.

Table of contents

Call Absolute Hearing Solutions
614-452-4280

About

Hearing aid parts and functions	8
Identify left and right hearing aid	10
MultiTool for handling batteries and cleaning	11

Start-up

Turn the hearing aid ON and OFF	12
When to replace the battery	13
How to replace the battery	14

Handling

Put on the hearing aid	16
Caring for the hearing aid	17
Flight mode	20

Options

Optional features and accessories	21
Change program	22
Change volume	23
Mute	24
Use the hearing aid with iPhone and iPad	25
Pair with iPhone	26

Wireless accessories	28
Other options	30
Tamper-resistant battery drawer	32

Tinnitus

Tinnitus SoundSupport™ (optional)	34
Guidelines for tinnitus sound generator users	35
Sound and volume adjustment	36
Limitation on use time	39


Warnings

Tinnitus SoundSupport warnings	42
General warnings	43

More info

Troubleshooting	52
Water and dust resistant (IP68)	54
Conditions of use	55
Technical information	56
Warranty	61
Mobile phone	62
Your individual hearing aid settings	64
Sound and LED indicators	66
Technical Data	70

Hearing aid parts and functions




Call Absolute Hearing Solutions
614-452-4280

The handling of BTE SP and BTE UP is the same. Therefore the illustrations in this booklet only show the BTE SP.

Identify left and right hearing aid

It is important to distinguish between the left and the right hearing aid as they may be programmed differently.

You can find left/right color indicators in the battery drawer. Indicators can also be found on the molds.




• A **BLUE** indicator marks the LEFT hearing aid

• A **RED** indicator marks the RIGHT hearing aid

Call Absolute Hearing Solutions
614-452-4280

MultiTool for handling batteries and cleaning

The MultiTool contains a magnet that makes it easier to replace the battery in the hearing aid. It also contains a brush and wire loop for cleaning and removing earwax. If you need a new MultiTool, please contact your hearing care professional.



IMPORTANT NOTICE

The MultiTool has a built-in magnet. Keep the MultiTool at least 12 inches away from credit cards and other magnetically sensitive devices.

Turn the hearing aid ON and OFF


The battery drawer is used to turn the hearing aid ON and OFF. To save battery life, make sure your hearing aid is turned OFF when you are not wearing it. When you turn ON the hearing aid, it will play a jingle. When you turn it OFF, it will play four descending tones.

If you wish to return to the standard settings of the hearing aid, simply open and close the battery drawer.

Optional LED

Two long green flashes indicate that the hearing aid is turned on.

Turn ON
Close the battery drawer with the battery in place.




Turn OFF
Open the battery drawer.




Call Absolute Hearing Solutions
614-452-4280

When to replace the battery

When it is time to replace the battery, you will hear three alternate tones repeated in moderate intervals until the battery runs out.

 **Three alternate tones***
= The battery is running low

 **Four descending tones**
= The battery has run out

Battery tip

Keep a spare battery on you at all times to ensure your hearing aid always works.


Optional LED

Continuous orange flashes indicate low battery.

* Bluetooth® will be turned off and it will not be possible to use wireless accessories.
Note: Batteries need to be replaced more often if you are streaming to your hearing aid.


How to replace the battery

1. Remove



Fully open the battery drawer. Remove the battery.


2. Uncover



Remove the sticky label from the + side of the new battery.

Tip:
Wait 2 minutes for the battery to draw air to ensure optimal functioning.

3. Insert



Insert the new battery into the battery drawer. Make sure the + side is facing up.

Insert it from above, NOT from the side.

Call Absolute Hearing Solutions
614-452-4280


4. Close



Close the battery drawer. The hearing aid will play a jingle through the earpiece.

Hold the earpiece close to your ear to hear the jingle.


Tip



The MultiTool can be used to change the battery. Use the magnetic end to remove and insert batteries.

The MultiTool is provided by your hearing care professional.

Put on the hearing aid



The mold is customized for you and fits your ear shape. The molds are unique for the left or the right ear.

Step 1



Gently pull your ear outwards and press the mold in the direction of the ear canal, twisting it slightly.

Step 2



Place the hearing aid behind your ear.


Call Absolute Hearing Solutions
614-452-4280

Caring for the hearing aid

When handling the hearing aid, hold it over a soft surface to avoid damage in case you drop it.

Clean the microphone openings

Use the brush of the MultiTool to carefully brush debris away from the openings. Carefully brush the surface around the openings.




Make sure that no parts of the MultiTool are squeezed into the microphone openings by force. This may damage the hearing aid.

IMPORTANT NOTICE

Use a soft, dry cloth to clean the hearing aid. The hearing aid must never be washed or immersed in water or other liquids.

Clean the mold

The mold should be cleaned regularly. Use a soft cloth to clean its surface. Use the MultiTool wire loop to clean the opening.



Replace the tube

The tubing between the mold and the hearing aid should be replaced if it becomes yellow or stiff. Consult your hearing care professional about this.

Call Absolute Hearing Solutions
614-452-4280

Wash the mold

Step 1




Make sure you are able to identify which mold is for which hearing aid before disconnecting them from the hearing aid.

Step 2



Disconnect the tube and mold from the hook. Hold the hook firmly when pulling off the tube. Otherwise the hearing aid may be damaged.

Step 3




Wash the mold and tube using mild soap. Rinse and dry completely before reconnecting to the hearing aid.

Flight mode

When flight mode is activated, Bluetooth® is turned off. The hearing aid will still be working. Pressing the push button on one hearing aid will activate flight mode on both hearing aids.

Optional LED

When flight mode is activated, the LED continuously flashes green, orange and orange.



- **To activate and deactivate**
Press button for at least seven seconds.
A short jingle confirms your action.

Opening and closing the battery drawer will also deactivate flight mode.
A short jingle confirms your action.

Call Absolute Hearing Solutions
614-452-4280

Optional features and accessories

The features and accessories described in the following pages are optional. Please contact your hearing care professional for more information about features and accessories.

If you experience difficult listening situations, a special program may be helpful. These are programmed by your hearing care professional.


Write down any hearing situations in which you may need help.

Change program

Your hearing aid can have up to four different programs. You will hear one to four tones when you change program depending on the program.

Optional LED

The number of green flashes indicates the selected program.



- Press the push button to switch between programs.

The program cycle switches one program forward, for example program 1 to 2 or program 4 to 1.


Call Absolute Hearing Solutions
614-452-4280

Change volume


The volume button allows you to adjust the volume. You will hear a beep when you increase or decrease the volume.

Optional LED

A green flash confirms a change of volume. A long green flash indicates that start-up volume has been reached. An orange flash will appear when you reach the minimum or maximum level.



- Press button to increase volume
- Press button to decrease volume



Mute

Use the mute function if you need to turn off the sounds temporarily, while wearing the hearing aid. The mute function only mutes the microphones on the hearing aid.

Optional LED

Continuous orange and green flashes indicate mute.



• **To mute the hearing aid**

Press either end of the volume button for at least four seconds.

To reactivate, press either end of the button briefly.

IMPORTANT NOTICE

Do not use the mute function as an off switch, as the hearing aid still draws current from the battery in this mode.

Call Absolute Hearing Solutions
614-452-4280

Use the hearing aid with iPhone and iPad

Your hearing aid is Made for iPhone® and allows for direct communication and control with iPhone, iPad® or iPod touch®. For assistance in using these products with your hearing aid, please contact your hearing care professional.




For information on compatibility, please visit www.oticon.com/support/compatibility

Use of the Made for Apple badge means that an accessory has been designed to connect specifically to the Apple product(s) identified in the badge, and has been certified by the developer to meet Apple performance standards. Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards.

Please note that use of this Accessory with iPod, iPhone, or iPad may affect wireless performance.


Pair with iPhone

1. Settings




Open your iPhone and go to "Settings". Make sure Bluetooth is on. Then choose "General".

2. General



On the "General" screen, choose "Accessibility".


3. Accessibility



On the "Accessibility" screen, choose "MFi Hearing Devices".


Call Absolute Hearing Solutions
614-452-4280

4. Prepare




Open and close the battery drawer on both hearing aids, and place them close to your iPhone. The hearing aids remain in pairing mode for 3 minutes.

5. Select



The iPhone will detect the hearing aids for pairing. Detected devices will appear in your iPhone list. Choose your hearing aids.

6. Confirm pairing



Confirm pairing. If you have two hearing aids, pairing confirmation is needed for each hearing aid.

Your hearing aids and iPhone will no longer be connected if either of them is turned off. Open and close the battery drawer of your hearing aids to reconnect them with your iPhone.

Wireless accessories

Call Absolute Hearing Solutions
614-452-4280

As an enhancement to your wireless hearing aids, a range of accessories is available. These can enable you to hear and communicate better in many everyday situations.

ConnectClip

When paired with ConnectClip, your hearing aids can be used as a hands-free headset. ConnectClip can also be used as a remote microphone.

TV Adapter 3.0

Streams sound directly from a TV or electronic audio device to your hearing aid.

Phone Adapter 2.0

Connects wirelessly to the hearing aids via ConnectClip allowing for use of landline phones.

Remote Control 3.0

Offers the ability to change program, adjust volume, or mute your hearing aid.

Oticon ON App

Intuitive and discreet control of your hearing aid. For iPhone, iPad, iPod touch, and Android™ devices.

For more information visit www.oticon.com/support or contact your hearing care professional.

Apple, the Apple logo, iPhone, iPad, and iPod touch are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc. Android, Google Play, and the Google Play logo are trademarks of Google LLC.

Other options



Telecoil program

Telecoil helps you hear better when using a telephone with a built-in loop, or when you are in buildings with teleloop systems such as theaters, churches, or lecture rooms. This symbol or a similar sign is shown wherever a teleloop has been installed.

Direct Audio Input (DAI)

Allows your hearing aid to receive signals directly from external sources such as TV, radio, music players, etc. The DAI adaptor is mounted to your hearing aid and via a cable connected to an external sound source. To attach a DAI adapter your HCP needs to replace the battery drawer.

Call Absolute Hearing Solutions
614-452-4280

FM

Allows you to receive speech or audio signals directly to your hearing aid from a dedicated transmitter. To attach an FM receiver or an FM adapter your HCP needs to replace the battery drawer.

CROS

Solution for people with an unaidable hearing loss in one ear. A transmitter on the poorer ear transmits sound to a hearing aid on the better ear.

IMPORTANT NOTICE

When the DAI is connected to an audio source plugged into a wall outlet, the audio source must comply with IEC 62368 or equivalent safety standards.

For more information, please contact your hearing care professional.

Tamper-resistant battery drawer

To keep the battery out of reach of infants, small children, and people with learning difficulties, a tamper-resistant battery drawer should be used.


Call Absolute Hearing Solutions
614-452-4280

Lock the battery drawer




Close the battery drawer completely and make sure it is locked.

Tool to unlock the battery drawer



Use this small tool provided by your hearing care professional.

Unlock the battery drawer



Insert the tip of the tool into the small hole at the back of the battery drawer.

1. Move it to the right and hold it.
2. Open the battery drawer.

IMPORTANT NOTICE

Avoid excessive force when opening the battery drawer in the locked position. Do not force the battery drawer beyond its fully opened position. Make sure to insert the battery correctly.

Please contact your hearing care professional if you suspect the tamper-resistant effectiveness or locking ability is damaged.

☐ **Tinnitus SoundSupport™** (optional)

Intended use of Tinnitus SoundSupport

Tinnitus SoundSupport is a tool intended to generate sounds to provide temporary relief for patients suffering from tinnitus as part of a tinnitus management program.

The target population is the adult population (over 18 years old).

Tinnitus SoundSupport is targeted to licensed hearing care professionals (audiologists, hearing aid specialists, or otolaryngologists) who are familiar with the evaluation and treatment of tinnitus and hearing loss. Fitting of Tinnitus SoundSupport must be done by a hearing care professional participating in a tinnitus management program.

Call Absolute Hearing Solutions
614-452-4280

Guidelines for tinnitus sound generator users

These instructions contain information about Tinnitus SoundSupport, which may have been enabled in your hearing aids by your hearing care professional.

Tinnitus SoundSupport is a tinnitus management device intended to generate sound of sufficient intensity and bandwidth to help manage tinnitus.

Your hearing care professional will also be able to offer the appropriate follow-up care. It is important to follow his/her advice and directions regarding such care.

Prescription use only

Good health practice requires that the person reporting tinnitus has a medical evaluation by a licensed ear physician before using a sound generator. The purpose of such an evaluation is to ensure that any medically treatable condition that may cause tinnitus is identified and treated prior to using a sound generator.

Sound and volume adjustment

Tinnitus SoundSupport is programmed by your hearing care professional to match your hearing loss and preferences for tinnitus relief. It offers a number of different sound options. Together with your hearing care professional, you can select your preferred sound(s).

Tinnitus SoundSupport programs

Together with your hearing care professional, you can decide for which programs you may want to have Tinnitus SoundSupport activated. The sound generator can be activated in up to four different programs.

Call Absolute Hearing Solutions
614-452-4280

Mute

If you are in a program for which Tinnitus SoundSupport is activated, the mute functionality will mute only the environmental sounds, and not the sound from Tinnitus SoundSupport. See chapter: Mute.

Volume adjustments with Tinnitus SoundSupport

When you select a hearing aid program for which Tinnitus SoundSupport is activated, your hearing care professional can only set the push button on your hearing aid to work as a volume control for the tinnitus relief sound.

Your hearing care professional will set the volume control for the sound generator in one of two ways:

A) How to change Tinnitus SoundSupport volume in each ear separately

To **increase** volume (on one hearing aid only), use a short press on the upper part of the push button repeatedly until desired level is reached. To **decrease** volume (on one hearing aid only), use a short press on the lower part of the push button repeatedly until desired level is reached.

B) How to change Tinnitus SoundSupport volume in both ears simultaneously

You can use one hearing aid to increase/decrease the sound in both hearing aids. When changing the volume in one hearing aid, the volume on the other hearing aid will follow.

To **increase** volume, use a short press on the upper part of the push button repeatedly.

To **decrease** volume, use a short press on the lower part of the push button repeatedly.

Call Absolute Hearing Solutions
614-452-4280

Limitation on use time

Daily use

The volume levels of Tinnitus SoundSupport can be set to a level which could lead to permanent hearing damage when used for a prolonged period of time. Your hearing care professional will advise you of the maximum amount of time per day you should use Tinnitus SoundSupport. It should never be used at uncomfortable levels.

See table "Tinnitus SoundSupport: Limitation on use" in section "Your individual hearing aid settings" at the end of this booklet to learn how many hours per day you can safely use the relief sound in your hearing aids.

Important information for hearing care professionals about Tinnitus SoundSupport

Call Absolute Hearing Solutions
614-452-4280

Device description

Tinnitus SoundSupport is a module function that can be enabled in the hearing aids by a hearing care professional.

Maximum wearing time

The wearing time of Tinnitus SoundSupport will decrease as you increase the level above 80 dB(A) SPL. The fitting software will automatically display a warning when the hearing aid exceeds 80 dB(A) SPL. See “Max wearing time indicator” next to the tinnitus fitting graph in the fitting software.

The volume control is deactivated

By default, the volume control for the sound generator is deactivated in the hearing aid. Risk of noise exposure increases when the volume control is activated.

If the volume control is activated

A warning may be displayed if you activate the tinnitus volume control in the “Buttons & Indicators” screen. This occurs if the relief sound can be listened to at levels that may cause hearing damage. The “Max wearing time” table in the fitting software displays the number of hours the patient can safely use Tinnitus SoundSupport.

- Note the max wearing time for each program for which Tinnitus SoundSupport is activated.
- Write those values in the table “Tinnitus SoundSupport: Limitation on use”, found in the back of this booklet.
- Instruct your patient accordingly.

Tinnitus SoundSupport warnings

If your hearing care professional has activated the sound generator Tinnitus SoundSupport, please pay attention to the following warnings.

There are some potential concerns associated with the use of any sound generated by a tinnitus management device. Among them are the potential for worsening of tinnitus and/or a possible change in hearing thresholds.

Should you experience or notice a change in hearing or tinnitus, or any dizziness, nausea, headaches, heart palpitations or possible skin irritation at the point of contact with the device, you should immediately discontinue use of the device and seek a consultation with a medical, audiologist, or other hearing care professional.

As with any device, misuse of the sound generator feature could present some potentially harmful effects. Care should be taken to prevent the unauthorized use and to keep the device out of the reach of children and pets.

Maximum wearing time

Always follow the maximum wearing time per day of the Tinnitus SoundSupport advised by your hearing care professional. Prolonged use may lead to worsening of your tinnitus or of your hearing loss.

Call Absolute Hearing Solutions
614-452-4280

General warnings

For your personal safety and to ensure correct usage, you should familiarize yourself fully with the following general warnings before using your hearing aid. Please note that a hearing aid will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions. Furthermore, note that in most cases, infrequent use of a hearing aid does not permit a user to attain its full benefits.

Consult your hearing care professional if you experience unexpected operations or events with your hearing aid.

Usage of hearing aids

Hearing aids should be used only as directed and adjusted by your hearing care professional. Misuse can result in sudden and permanent hearing loss.

Never allow others to wear your hearing aid as incorrect usage could cause permanent damage to their hearing.

Choking hazards & risk of swallowing batteries and other small parts

Hearing aids, their parts, and batteries should be kept out of reach of children and anyone who might swallow these items or otherwise cause injury to themselves.

Batteries have occasionally been mistaken for pills. Therefore, check your medicine carefully before swallowing any pills.

Most hearing aids can be supplied with a tamper-resistant battery drawer upon request. This is strongly recommended for infants, small children, and people with learning difficulties.

Children younger than 36 months must always use a tamper-resistant battery drawer. Please talk to your hearing care professional about the availability of this option.

General warnings

If a battery or hearing aid is swallowed, see a doctor immediately and contact the National Poison Center at 1-800-222-1222 or National Battery Ingestion Hotline at 1-800-498-8666.

Battery use

Always use batteries recommended by your hearing care professional. Low quality batteries may leak and cause bodily harm.

Never attempt to recharge your batteries, and never dispose of batteries by burning them. There is a risk that the batteries will explode.

Use in explosive environments

- only applies to BTE UP hearing aid

The hearing aid is safe to use under normal usage conditions. The BTE UP hearing aid with size 675 battery has not been tested for compliance with international standards concerning explosive environments.

Therefore, do not use BTE UP in environments with danger of explosions e.g. mines, oxygen rich environments or areas where flammable anaesthetics are handled.

Dysfunction

Be aware of the possibility that your hearing aid may stop working without notice. Keep this in mind when you depend on warning sounds (e.g. when you are in traffic). The hearing aids may stop functioning, for instance if the batteries have expired or if the tubing is blocked by moisture or earwax.

Active implants

The hearing aid has been thoroughly tested and characterized for human health according to international standards for human exposure (Specific Absorption Ratio - SAR), induced electromagnetic power and voltages into the human body.

Call Absolute Hearing Solutions
614-452-4280

The exposure values are well below international accepted safety limits for SAR, induced electromagnetic power and voltages into the human body defined in the standards for human health and coexistence with active medical implants, such as pacemakers and heart defibrillators.

If you have an active brain implant, please contact the manufacturer of your implantable device for information about the risk of disturbance.

The MultiTool which has a built-in magnet should be kept at least 12 inches away from the implant, e.g. do not carry it in your breast pocket.

In general, please follow the guidelines recommended by the manufacturers of implantable defibrillators and pacemakers on use with magnets.

Cochlear implants

If you are using a cochlear implant (CI) on one ear and a hearing aid on the other ear, make sure to always keep your CI antenna and magnets at a safe distance from your hearing aid. The magnetic field from the CI magnets may permanently damage the speaker unit in your hearing aid.

Never place the devices close together on a table e.g. when cleaning or changing batteries. Do not carry the CI system and the hearing aid together in the same box.

General warnings

X-ray/CT/MR/PET scanning, and electrotherapy

Remove your hearing aid before X-ray/CT/MR/PET scanning, electrotherapy, surgery, etc. as your hearing aid may be damaged when exposed to strong electromagnetic fields.

The hearing aid must never be exposed to extreme heat, e.g. left inside a parked car in the sun.

The hearing aid must not be dried in microwave ovens or other ovens.

The chemicals in cosmetics, hairspray, perfume, aftershave lotion, sunscreen lotion, and insect repellent can damage the hearing aid. Always remove your hearing aid before applying such products and allow time to dry before use.

Power instrument

Special care should be exercised when selecting, fitting and using a hearing aid when the maximum sound pressure capability exceeds 132 dB SPL (IEC 711), as there may be risk of impairing the remaining hearing of the hearing aid user.

Possible side effects

Hearing aids and earpieces may cause an accelerated accumulation of earwax.

The non-allergenic materials used in hearing aids may in rare cases cause a skin irritation or other side effects.

Please consult a physician if these conditions occur.

Call Absolute Hearing Solutions
614-452-4280

Interference

The hearing aid has been thoroughly tested for interference, in accordance with the most stringent international standards. However, interference between the hearing aid and other devices (e.g. some mobile telephones, citizens band systems, shop alarm systems and other devices) may occur. If this occurs, increase the distance between the hearing aid and the interfering device.

Use on aircraft

Your hearing aid has Bluetooth. On board an aircraft, flight mode must be activated, unless Bluetooth is permitted by the flight personnel.

Connection to external equipment

The safety of the use of hearing aids when connected to external equipment (with an auxiliary input cable, USB cable or directly) is determined by the external equipment. When connected to external equipment plugged into a wall outlet, this equipment must comply with IEC 62368 or equivalent safety standards.

Incompatible non-supplied accessories

Only use accessories, transducers or cables supplied by the manufacturer of this device. Incompatible accessories may result in increased emission or decreased immunity.

Modification of hearing aids is not allowed

Changes or modifications not expressly approved by the manufacturer will void the warranty of the equipment.

General warnings

Warning to hearing care professional

A hearing care professional should advise a prospective hearing aid user to consult immediately with a licensed physician (preferably an ear specialist) before dispensing a hearing aid if the hearing care professional determines through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions:

- (i) Visible congenital or traumatic deformity of the ear.
- (ii) History of active drainage from the ear within the previous 90 days.
- (iii) History of sudden or rapidly progressive hearing loss within the previous 90 days.
- (iv) Acute or chronic dizziness.
- (v) Unilateral hearing loss of sudden or recent onset within the previous 90 days.

- (vi) Audiometric air-bone gap equal to or greater than 15 decibels at 500 Hertz (Hz), 1,000 Hz, and 2,000 Hz.
- (vii) Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- (viii) Pain or discomfort in the ear.

Call Absolute Hearing Solutions
614-452-4280

Special care should be exercised in selecting and fitting a hearing aid whose maximum sound pressure capability exceeds 132 dB SPL as there may be risk of impairing the remaining hearing of the hearing aid user.

Important notice for prospective hearing aid users

Good health practice requires that a person with a hearing loss have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before purchasing a hearing aid. Licensed physicians who specialize in diseases of the ear are often referred to as Otolaryngologists, Otologists or Otorhinolaryngologists. The purpose of medical evaluation is to ensure that all medically treatable conditions that may affect hearing are identified and treated before the hearing aid is purchased. Following the medical evaluation, the physician will give you a written statement that states that your hearing loss has been medically

evaluated and that you may be considered a candidate for a hearing aid. The physician will refer you to an audiologist or a hearing aid dispenser, as appropriate, for a hearing aid evaluation.

The audiologist or hearing care professional will conduct a hearing aid evaluation to assess your ability to hear with and without a hearing aid. The hearing aid evaluation will enable the audiologist or dispenser to select and fit a hearing aid to your individual needs. If you have reservations about your ability to adapt to amplification, you should inquire about the availability of a trial, rental or purchase-option program. Many hearing care professionals now offer programs that permit you to wear a hearing aid for a period of time for a nominal fee, after which you may decide if you want to purchase the hearing aid. Federal law limits the sale of hearing aids to those individuals who have obtained a medical evaluation from a licensed physician.

Continues on next page

General warnings

Federal law permits a fully informed adult to sign a waiver statement declining the medical evaluation for religious or personal beliefs that preclude consultation with a physician. The exercise of such a waiver is not in your best health interest and its use is strongly discouraged. A hearing aid will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions. A hearing aid is only part of hearing rehabilitation and may need to be supplemented by auditory training and lip reading.

Children with hearing loss

In addition to seeing a physician for medical evaluation, a child with a hearing loss should be directed to an audiologist for evaluation and rehabilitation, since hearing loss may cause problems in language development and educational and social growth of a child. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of a child with a hearing loss. If the user is an infant, small child, or person of mental incapacity, it is recommended that the hearing aid be modified with a tamper-resistant battery compartment.

Call Absolute Hearing Solutions
614-452-4280

Troubleshooting

Call Absolute Hearing Solutions
614-452-4280

Symptom	Possible causes	Solutions
No sound	Dead battery	Replace the battery
	Clogged mold	Clean mold
	Microphones muted	Unmute microphones
Intermittent or reduced sound	Clogged sound outlet	Clean mold
	Moisture	Wipe battery with a dry cloth
	Dead battery	Replace the battery
Squealing noise	Mold inserted incorrectly	Re-insert the mold
	Earwax accumulated in ear canal	Have ear canal examined by your doctor
	Ear has grown	Contact your hearing care professional
Beeping/orange flashes (optional)	If your hearing aid plays 8 beeps, 4 times consecutively, and/or repeatedly flashes orange 4 times with small pauses, your hearing aid needs a microphone service check	Contact your hearing care professional
Pairing issue with Apple device	Bluetooth connection failed	1) Unpair your hearing aid (Settings→General→Accessibility→Hearing Devices→Devices→Forget this device). 2) Turn Bluetooth off and on again. 3) Open and close battery drawer on hearing aid. 4) Re-pair hearing aid (see section: Pair with iPhone)
	Only one hearing aid paired	

If none of the above solutions work, consult your hearing care professional.

Water and dust resistant (IP68)

Your hearing aid is dust tight and protected against ingress of water, which means it is designed to be worn in all daily life situations. Therefore, you do not have to worry about sweat or getting wet in the rain. Should your hearing aid come into contact with water and stop working, please follow these guidelines:

- 1. Gently wipe off any water
- 2. Open the battery drawer, remove the battery, and gently wipe off any water in the battery drawer
- 3. Let the hearing aid dry with the battery drawer left open for approximately 30 minutes
- 4. Insert a new battery

Call Absolute Hearing Solutions
614-452-4280

Conditions of use

Operating conditions	Temperature: +34 °F to +104 °F Relative humidity: 5% to 93%, non-condensing
Storage and transport conditions	Temperature and humidity shall not exceed the below limits for extended periods during transport and storage: Temperature: -13 °F to +140 °F Relative humidity: 5% to 93%, non-condensing

IMPORTANT NOTICE

Do not wear your hearing aid while showering or participating in water activities. Do not immerse your hearing aid in water or other liquids.

Mobile phone

Some hearing aid users have reported a buzzing sound in their hearing aid when they are using mobile phones, indicating that the mobile phone and hearing aid may not be compatible.

The ANSI C63.19 standard determines the prediction of compatibility between a specific hearing aid and a mobile phone by: adding the numerical value of the rating for the hearing aid immunity to the numerical value of the rating for the mobile phone emissions. A sum of 4 would indicate that the combination of wireless device and hearing aid is usable; a combined rating that equals at least 5 would provide normal use; a combined rating of 6 or greater would indicate excellent performance.

Whereas all hearing aids have acoustic coupling, only the larger hearing aids have the physical space for telecoil (inductive) coupling. These two types of coupling have different rating scales (M1-M4 for acoustic coupling and T1-T4 for telecoil coupling, respectively) and both ratings are therefore relevant when predicting the compatibility of a particular hearing aid.

For a hearing aid with both acoustic coupling and telecoil coupling with a rating of M4/T2 and with a telephone rating of M3/T3), the combined rating is 7 (M4 + M3) for the acoustic coupling and 5 (T2 + T3) for the telecoil coupling. According to the guidelines given above, both types of coupling will thereby be

Call Absolute Hearing Solutions
614-452-4280

acceptable, with the acoustic coupling indicating excellent performance and the telecoil coupling indicating normal use.

The above equipment performance measurements, categories and system classifications are based upon the best information available, but it cannot be guaranteed that all users will be satisfied.

The immunity of BTE SP and BTE UP is at least M2/T2.

The equipment performance measurements, categories and system classifications are based upon the best information available but cannot guarantee that all users will be satisfied.

IMPORTANT NOTICE

The performance of individual hearing aids may vary with individual mobile phones. Therefore, please try this hearing aid with your mobile phone or, if you are purchasing a new phone, be sure to try it with your hearing aid prior to purchase. For additional guidance, please ask your mobile phone provider for the booklet entitled "Hearing Aid Compatibility with Digital Wireless Cell Phones."

Technical information

The hearing aid contains two radio technologies, which are described below:

The hearing aid contains a radio transceiver that uses short-range magnetic induction technology working at 3.84 MHz. The magnetic field strength of the transmitter is very weak and is always below 15 nW (typically below -40 dBμA/m at a 10 meter distance).

The hearing aid also contains a radio transceiver using Bluetooth Low Energy (BLE) and a proprietary short-range radio technology both working at ISM band 2.4 GHz. The radio transmitter is weak and always below 3 mW equal to 4.8 dBm in total radiated power.

The hearing aid complies with international standards concerning electromagnetic compatibility and human exposure.

Due to the limited space available on the hearing aid, relevant approval markings can be found in this booklet.

Additional information can be found in the "Technical Data sheet" on www.oticon.com

Call Absolute Hearing Solutions
614-452-4280

The manufacturer declares that this hearing aid is in compliance with the essential requirements and other relevant provisions of Directive 2014/53/EU.

Declaration of Conformity is available from the manufacturer.



Manufactured by:
Oticon A/S
Kongebakken 9
DK-2765 Smørum
Denmark
www.oticon.global

CE 0543









Waste from electronic
equipment must be
handled according to
local regulations.




 **Bluetooth®**







IP68

Description of symbols used in this booklet	
	Warnings Text marked with a warning symbol must be read before using the device.
	Manufacturer The device is produced by the manufacturer whose name and address are stated next to the symbol. Indicates the medical device manufacturer, as defined in EU Directives 90/385/EEC, 93/42/EEC and 98/79/EC.
	CE mark The device complies with Medical Device Directive 93/42/EEC. The four-digit number indicates the identification of the notified body.
	Electronic waste (WEEE) Recycle hearing aids, accessories or batteries according to local regulations. Hearing aid users can also return electronic waste to their hearing care professional for disposal. Electronic equipment covered by Directive 2012/19/EU on waste and electrical equipment (WEEE).
	Regulatory Compliance Mark (RCM) The device complies with electrical safety, EMC and radio spectrum requirements for devices supplied to the Australian or New Zealand market.
	IP code Indicates the class of protections against harmful ingress of water and particulate matter according to EN 60529. IP6X indicates total dust protection. IPX8 indicates the protection against the effects of continuous immersion in water.

Call Absolute Hearing Solutions
614-452-4280

	Bluetooth logo Registered trademark of Bluetooth SIG, Inc. where any use of such requires a license.
	Made for Apple badges Indicates that the device is compatible with iPhone, iPad and iPod touch.
	Hearing loop This logo incorporates the universal symbol for hearing assistance. The "T" signifies that a hearing loop is installed.

Description of symbols used on the regulatory packaging label	
	Keep dry Indicates a medical device that needs to be protected from moisture.
	Caution symbol Consult instructions for use for warnings and cautions.
	Catalog number Indicates the manufacturer's catalog number so that the medical device can be identified.
	Serial number Indicates the manufacturer's serial number so that a specific medical device can be identified.

International warranty

Your hearing aid is covered by an international limited warranty issued by the manufacturer for a period of 12 months from the date of delivery. This limited warranty covers manufacturing and material defects in the hearing aid itself, but not in accessories such as batteries, tubing, speakers, earpieces and filters, etc. Problems resulting from improper or incorrect handling or care, excessive use, accidents, repairs made by an unauthorized party, exposure to corrosive conditions, physical changes in your ear, damage due to foreign objects entering the device, or incorrect adjustments are NOT covered by the limited warranty and may void it. The above warranty does not affect any legal rights that you might have under applicable national legislation governing sale of consumer goods. Your hearing care professional may have

issued a warranty that goes beyond the clauses of this limited warranty. Please consult him/her for further information.

If you need service

Take your hearing aid to your hearing care professional, who may be able to sort out minor problems and adjustments immediately.

Call Absolute Hearing Solutions
614-452-4280

Warranty Certificate

Name of owner: _____

Hearing care professional: _____

Hearing care professional's address: _____

Hearing care professional's phone: _____

Purchase date: _____

Warranty period: _____ Month: _____

Model left: _____ Serial no.: _____

Model right: _____ Serial no.: _____

Your individual hearing aid settings

To be filled out by your hearing care professional.

Call Absolute Hearing Solutions
614-452-4280











Settings overview for your hearing aid				
Left			Right	
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Change volume	<input type="checkbox"/> Yes	<input type="checkbox"/> No
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Change program	<input type="checkbox"/> Yes	<input type="checkbox"/> No
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Mute	<input type="checkbox"/> Yes	<input type="checkbox"/> No
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Tinnitus SoundSupport	<input type="checkbox"/> Yes	<input type="checkbox"/> No
Volume control indicators				
<input type="checkbox"/> On	<input type="checkbox"/> Off	Beeps at min/max volume	<input type="checkbox"/> On	<input type="checkbox"/> Off
<input type="checkbox"/> On	<input type="checkbox"/> Off	Beeps when changing volume	<input type="checkbox"/> On	<input type="checkbox"/> Off
<input type="checkbox"/> On	<input type="checkbox"/> Off	Beeps at start-up volume	<input type="checkbox"/> On	<input type="checkbox"/> Off
Battery indicators				
<input type="checkbox"/> On	<input type="checkbox"/> Off	Low battery warning	<input type="checkbox"/> On	<input type="checkbox"/> Off

Tinnitus SoundSupport: Limitation on use			
<input type="checkbox"/>	No limitation on use		
	Program	Start-up volume (Tinnitus)	Max volume (Tinnitus)
<input type="checkbox"/>	1	Max _____ hours per day	Max _____ hours per day
<input type="checkbox"/>	2	Max _____ hours per day	Max _____ hours per day
<input type="checkbox"/>	3	Max _____ hours per day	Max _____ hours per day
<input type="checkbox"/>	4	Max _____ hours per day	Max _____ hours per day

Sound and LED indicators

Different sounds and LED lights indicate the hearing aid status. The different indicators are listed on the following pages.








Your hearing care professional can set sound and LED indicators to match your preferences.

Program	<input type="checkbox"/> Sound	<input type="checkbox"/> LED ¹⁾	When to use
1	1 tone		
2	2 tones	 	
3	3 tones	  	
4	4 tones	   	

 Green flash










1) LED continuous or repeated three times with small pauses

Call Absolute Hearing
614-452-4280

Sound	Sound	LED	LED comments
On	<input type="checkbox"/> Jingle	<input type="checkbox"/>  	Shown one time
Volume	Sound	LED	
Start-up volume	<input type="checkbox"/> 2 beeps	<input type="checkbox"/> 	Shown one time
Minimum/maximum volume	<input type="checkbox"/> 3 beeps	<input type="checkbox"/> 	
Volume up/down	<input type="checkbox"/> 1 beep	<input type="checkbox"/> 	
Mute	<input type="checkbox"/> Descending tone	<input type="checkbox"/>  	Continuous or repeated three times
Unmute	<input type="checkbox"/> Ascending tone		

 Long green flash  Green flash  Orange flash












Continues on next page

Accessories	<input type="checkbox"/> Sound	<input type="checkbox"/> LED	LED comments
FM and wireless programs ¹⁾	2 different tones	 	Continuous or repeated three times
FM and wireless programs ²⁾	2 different tones	 	
Flight mode	Sound	<input type="checkbox"/> LED	LED comments
Flight mode active	Short jingle	  	Continuous or repeated three times
Flight mode inactive	Short jingle	  ³⁾	

 Long green flash  Green flash  Orange flash

- 1) DAI/FM + hearing aid microphone, TV Adapter
 2) DAI/FM signal only, ConnectClip remote microphone
 3) Only available when three-time repetition is selected

Call Absolute Hearing Solutions
 614-452-4280

Solutions	Sound	LED	LED comments
Low battery	<input type="checkbox"/> 3 alternate tones	<input type="checkbox"/>       	Continuously flashing
Battery shut down	4 descending tones		
Microphone service check needed	8 beeps repeated 4 times	<input type="checkbox"/>    	Repeated four times

 Orange flash  Long orange flash

Technical Data

BTE SP


Can Absolute Hearing Solutions

614-452-4280

110 **2CC Coupler** measured according to American National Standard ANSI S3.22-2014 and ANSI S3.55-2014/Part 5. Supply voltage: Battery Zinc Air 1.4 Volt

110

Oticon Xceed



0 dB SPL ref. 20 mPa

Oticon Xceed

Peak OSPL90	139 dB SPL
HF Average OSPL90	130 dB SPL
Peak Full-on Gain	79 dB
HF Average Full-on Gain	70 dB
Reference Test Gain	53 dB
Frequency Range	100-6100 Hz
Total Harmonic Distortion 500 Hz	4 %
Total Harmonic Distortion 800 Hz	< 2%
Total Harmonic Distortion 1600 Hz	< 2%
Battery Consumption (Typical)	2.5 mA
Equivalent Input Noise Level (omni/dir)	19/34 dB SPL
HF Average SPLITS (left/right ear)	115/115 dB SPL
Attack Time	< 5
Release Time	< 90

Technical Data

BTE UP

Call Absolute Hearing Solutions

614-452-4280

120

2CC Coupler measured according to American National Standard ANSI S3.22-2014 and ANSI S3.55-2014/Part 5. Supply voltage: Battery Zinc Air 1.4 Volt

120


Oticon Xceed

OSPL90 - Output Sound

Pressure Level

Input: 90 dB SPL.

Technical setting: A0




Full-on Gain

Input: 50 dB SPL.

Technical setting: A0

— Acoustic input: 60 dB SPL
 --- Magnetic input: 31.6 mA/m



0 dB SPL ref. 20 mPa

Oticon Xceed

Peak OSPL90	142 dB SPL
HF Average OSPL90	130 dB SPL
Peak Full-on Gain	83 dB
HF Average Full-on Gain	69 dB
Reference Test Gain	53 dB
Frequency Range	100-5300 Hz
Total Harmonic Distortion 500 Hz	9 %
Total Harmonic Distortion 800 Hz	< 2%
Total Harmonic Distortion 1600 Hz	3 %
Battery Consumption (Typical)	4.1 mA
Equivalent Input Noise Level (omni/dir)	23/38 dB SPL
HF Average SPLITS (left/right ear)	112/112 dB SPL
Attack Time	< 5
Release Time	< 90

Technical Data

BTE SP

Can Absolute Hearing Solutions

614-452-4280

110

2CC Coupler measured according to American National Standard ANSI S3.22-2014 and ANSI S3.55-2014/Part 5. Supply voltage: Battery Zinc Air 1.4 Volt

110


Oticon Xceed Play

OSPL90 - Output Sound

Pressure Level

Input: 90 dB SPL.

Technical setting: A0




Full-on Gain

Input: 50 dB SPL.

Technical setting: A0

— Acoustic input: 60 dB SPL
 --- Magnetic input: 31.6 mA/m



0 dB SPL ref. 20 mPa

Oticon Xceed Play

Peak OSPL90	139 dB SPL
HF Average OSPL90	130 dB SPL
Peak Full-on Gain	79 dB
HF Average Full-on Gain	70 dB
Reference Test Gain	53 dB
Frequency Range	100-6100 Hz
Total Harmonic Distortion 500 Hz	4 %
Total Harmonic Distortion 800 Hz	< 2%
Total Harmonic Distortion 1600 Hz	< 2%
Battery Consumption (Typical)	2.5 mA
Equivalent Input Noise Level (omni/dir)	19/34 dB SPL
HF Average SPLITS (left/right ear)	115/115 dB SPL
Attack Time	< 5
Release Time	< 90

Technical Data

BTEUP

Call Absolute Hearing Solutions


614-452-4280

120

2CC Coupler measured according to American National Standard ANSI S3.22-2014 and ANSI S3.55-2014/Part 5. Supply voltage: Battery Zinc Air 1.4 Volt

120

Oticon Xceed Play



208033US / 2019.07.12

0 dB SPL ref. 20 mPa

Oticon Xceed Play

Peak OSPL90	142 dB SPL
HF Average OSPL90	130 dB SPL
Peak Full-on Gain	83 dB
HF Average Full-on Gain	69 dB
Reference Test Gain	53 dB
Frequency Range	100-5300 Hz
Total Harmonic Distortion 500 Hz	9 %
Total Harmonic Distortion 800 Hz	< 2%
Total Harmonic Distortion 1600 Hz	3 %
Battery Consumption (Typical)	4.1 mA
Equivalent Input Noise Level (omni/dir)	23/38 dB SPL
HF Average SPLITS (left/right ear)	112/112 dB SPL
Attack Time	< 5
Release Time	< 90

208033US / 2019.07.12

Call Absolute Hearing Solutions
614-452-4280



Absolute Hearing Solutions LLC
absolutehearing@att.net

750 Cross Pointe Road Suite F
Gahanna, Ohio 43230

Phone: 614-452-4280
Toll Free: 866-603-2159
Fax: 614-577-0481
www.absolutehearingolutions.com



0000208033000001

oticon
PEOPLE FIRST